

## PyroPure® Recommended Service Schedule



### ONE-YEAR MAINTENANCE

- Visual Inspection
- Float Ball Replacement
- Replace Cooling Water Strainer
- Visual Inspection of Feedwater Conductivity Probe (If Equipped)
- Visual Inspection of Distillate Conductivity Probe (If Equipped)
- Rebuild Steam Traps
- Replace Feedwater Valve Diaphragms
- Replace Sample Valve Diaphragm (If Equipped)
- Replace Check Valves
- Replace Instrument Air Inlet Filter
- Inspect/Adjust Steam Control Valve
- Inspect/Adjust Distillate Temperature Control Valve (If Equipped)
- Startup
- Leak Check
- Alarm Function Test
- Adjust and Tune as Necessary
- Document Current Settings
- Provide a Service Report



### TWO-YEAR MAINTENANCE

- Includes Everything from One-Year Maintenance
- Replace Distillate to Tank Valve Diaphragm (If Equipped)
- Replace Distillate to Drain Valve Diaphragm (If Equipped)
- Replace Evaporator Seals
- Replace Column Seals (If Equipped)
- Replace Rupture Disc
- Replace Burst Sensor
- Adjust/Function Test